Pycnometers acc. to Gay Lussac, adjusted



Pycnometers acc. to Gay Lussac, adjusted

Gay Lussac pycnometers acc. to DIN ISO 3507 and in addition to standard, adjusted, stopper NS 10/19 with capillary, stopper top ground and polished, adjusted to ‘In’ (with single certificate), accuracy: 0.001 cm³, material: borosilicate glass 3.3

Nominal capacity (cm³)	DIN ISO	Pack (qty.)	Cat. No.
5	–	1	5 6302 10
10	3507	1	5 6302 13
25	3507	1	5 6302 22
50	3507	1	5 6302 28
100	–	1	5 6302 37

Generated on 2026-11-17